

IV Assembly Instructions

1



b + c

Snap membrane onto valve until membrane lies completely flat.

2



+



a

b c

Push assembled valve and membrane onto breastshield. Ensure sidewise position of valve as indicated on picture.

3



+



a b c

d

Screw bottle into breastshield.

4



+

i

a b c d

Align the raised edge on the breastshield back with the ridge on the underside of the battery cover and press the motor firmly onto the shield. Lock the motor onto the breastshield by rotating clockwise until the stop.